

Observation from Ambulance.

18/11/2023

>) Defibrillator :-

⇒ Pacemaker Software not loaded.

as per our offer this was included DFBLIIPACE Pacemaker module complete of one set of defi pads and adapter cable

⇒ 10/12 lead ECG cable not available.

SET12 leads module included (specification working with 10 leads patient cable only)

⇒ Electrical not working (because no socket for power supply).

>) Portable Ventilator.

as we mentioned initially in our offer that we will provide medical devices which will be available in our stocks

⇒ Only one language available (not 20 language)

⇒ User manual is not in English (It is in Arabic.)

⇒ Available hose DIN type. (not matching our Ambulance, because it is BS type connector is thinner (BS type connector is thinner than DIN type.)

⇒ Wall bracket need to fix. (for holding vent.)

>)

Suction machine :-

⇒ Electrical not working (because no socket for power supply)

⇒ Suction machine fixed holder Area need change.

Uttre down side (Now after filling waste fluid not able to take out)

Note: we provided the standard Suction with bracket to hold this Suction Device.

>)

Nebulizer please be advise that we informed you about this items, the Myneb Brand is not available and we sent the catalog of alternative brand which is Trister see attached screenshot of our emails

⇒ Myneb brand not provided.

⇒ They provide Trister

⇒ Trister is connected in USD type.

>) Need to Add 3 ^{extremes} (Ambulance back side no USD Port, power supply)